
Midmark M11 - v1.0.0
Total Cycles: 4990

Name: RITTER 1
Sterilizer ID: V2178198
10/11/2024 09:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.6 F 0.0
2:00 162.0 F 0.5
4:00 170.3 F 2.3
6:00 188.5 F 6.0
8:00 206.9 F 10.1
10:00 224.2 F 7.9
12:00 229.8 F 8.8
14:00 239.9 F 13.5
16:00 238.7 F 10.5
18:00 234.8 F 7.9
20:00 248.5 F 14.3
22:00 259.4 F 20.3
24:00 269.5 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.1
1:00 272.1 F 29.0
2:00 271.6 F 28.4
3:00 272.6 F 29.0
4:00 271.3 F 28.6

Min 270.6 F 28.0
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 196.1 F 0.0
10:00 215.1 F 0.0
15:00 218.1 F 0.0
20:00 242.0 F 0.0
25:00 192.9 F 0.0
30:00 187.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:59
STERILIZING: 4:00
VENTING: 3:44
DRYING: 30:00
TOTAL CYCLE: 01:03:37
CYCLE COMPLETE

10/11/2024 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____