
Midmark M11 - v1.0.0
Total Cycles: 3692

Name: RITTER 1
Sterilizer ID: V2178198
08/28/2023 02:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 183.2 F 0.0
2:00 174.8 F 3.3
4:00 189.6 F 7.5
6:00 207.6 F 12.2
8:00 223.4 F 6.2
10:00 236.1 F 11.2
12:00 233.2 F 7.9
14:00 244.1 F 12.1
16:00 244.7 F 12.4
18:00 256.8 F 18.9
20:00 267.8 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.0 F 28.4
2:00 272.5 F 28.8
3:00 271.5 F 28.3
4:00 272.0 F 28.4

Min 271.3 F 28.0
Max 272.6 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.5 F 0.0
5:00 233.0 F 0.0
10:00 225.3 F 0.0
15:00 225.5 F 0.0
20:00 223.5 F 0.0
25:00 221.3 F 0.0
30:00 221.8 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 21:16
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:18
CYCLE COMPLETE

08/28/2023 03:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____