
Midmark M11 - v1.0.5
Total Cycles: 841

Name: RITTER2
Sterilizer ID: V2313228
02/28/2022 07:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.3 F 0.0
2:00 75.2 F 0.0
4:00 104.1 F 0.0
6:00 111.2 F 0.1
8:00 144.1 F 1.9
10:00 178.0 F 5.0
12:00 191.8 F 9.1
14:00 207.7 F 13.4
16:00 220.6 F 7.0
18:00 233.6 F 11.2
20:00 231.3 F 7.6
22:00 243.4 F 13.3
24:00 242.6 F 11.8
26:00 254.8 F 18.2
28:00 265.9 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.7
1:00 272.8 F 29.5
2:00 271.5 F 28.6
3:00 272.4 F 29.6
4:00 272.0 F 28.9

Min 271.2 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.3 F 0.0
5:00 207.0 F 0.0
10:00 213.2 F 0.0
15:00 194.6 F 0.0
20:00 196.9 F 0.0
25:00 213.6 F 0.0
30:00 212.3 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 29:46
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:07:40
CYCLE COMPLETE

02/28/2022 08:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____