
Midmark M11 - v1.0.5
Total Cycles: 1135

Name: RITTER2
Sterilizer ID: V2313228
05/30/2022 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.8 F 0.0
2:00 148.0 F 0.9
4:00 161.6 F 3.8
6:00 184.5 F 8.0
8:00 210.1 F 12.2
10:00 220.3 F 6.0
12:00 231.0 F 10.7
14:00 231.8 F 7.4
16:00 242.3 F 12.8
18:00 241.4 F 11.4
20:00 254.0 F 17.9
22:00 264.9 F 24.5
24:00 271.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.5
1:00 272.8 F 30.0
2:00 272.0 F 29.8
3:00 272.7 F 29.7
4:00 272.6 F 30.2

Min 271.8 F 29.4
Max 273.0 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 208.9 F 0.0
10:00 237.4 F 0.0
15:00 229.8 F 0.0
20:00 226.7 F 0.0
25:00 227.9 F 0.0
30:00 227.7 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:33
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:02:31
CYCLE COMPLETE

05/30/2022 09:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____