
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000268
8:14 01/10/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:51
PRESSURIZING 3:53
STERILIZING 5:23
136.4 C 232kPa 5:23
136.2 C 232kPa 5:53
136.3 C 233kPa 6:23
136.2 C 230kPa 6:53
136.3 C 233kPa 7:23
136.3 C 232kPa 7:53
136.3 C 231kPa 8:23
136.4 C 233kPa 8:53
136.4 C 233kPa 8:53
VENTING 8:53
AIR DRYING 10:01
CYCLE COMPLETE 25:01

Digital Signature #
F358DF752FC9E918
