
Midmark M11 - v1.0.0
Total Cycles: 1926

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2022 01:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.6 F 0.0
2:00 167.5 F 3.0
4:00 183.7 F 6.9
6:00 205.0 F 11.7
8:00 217.8 F 5.3
10:00 230.5 F 10.1
12:00 233.8 F 8.2
14:00 242.6 F 12.3
16:00 240.9 F 10.6
18:00 254.4 F 17.4
20:00 265.8 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.8 F 28.8
2:00 271.8 F 28.1
3:00 272.3 F 28.8
4:00 272.6 F 28.7

Min 271.5 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 202.9 F 0.0
10:00 208.1 F 0.0
15:00 205.7 F 0.0
20:00 202.4 F 0.0
25:00 203.6 F 0.0
30:00 203.5 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:59
CYCLE COMPLETE

05/10/2022 02:08 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____