
Midmark M11 - v1.0.0
Total Cycles: 1104

Name: RITTER 1
Sterilizer ID: V2178198
08/20/2021 09:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.3 F 0.0
2:00 144.2 F 0.0
4:00 164.3 F 0.1
6:00 190.7 F 2.4
8:00 205.4 F 6.0
10:00 219.8 F 10.0
12:00 230.4 F 8.5
14:00 232.1 F 9.3
16:00 241.1 F 10.5
18:00 241.1 F 11.9
20:00 239.8 F 10.3
22:00 252.5 F 16.7
24:00 263.7 F 23.3
26:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.2
2:00 271.5 F 28.5
3:00 271.9 F 29.2
4:00 272.3 F 29.0

Min 271.2 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.2 F 0.0
5:00 215.1 F 0.0
10:00 219.4 F 0.0
15:00 215.5 F 0.0
20:00 213.9 F 0.0
25:00 216.4 F 0.0
30:00 218.7 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 26:10
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:04:17
CYCLE COMPLETE

08/20/2021 10:12 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____