
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004718
11:32 15/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:31
STERILIZING 3:32
136.4 C 220kPa 3:32
136.4 C 225kPa 4:02
136.3 C 223kPa 4:32
136.4 C 223kPa 5:02
136.4 C 219kPa 5:32
136.3 C 224kPa 6:02
136.4 C 224kPa 6:32
136.3 C 218kPa 7:02
136.3 C 218kPa 7:02
VENTING 7:02
AIR DRYING 7:54
CYCLE COMPLETE 8:02

Digital Signature #
EC9EBA9AAEDE67A6
