
Midmark M11 - v1.0.5
Total Cycles: 596

Name: RITTER2
Sterilizer ID: V2313228
12/01/2021 02:21 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.3 F 0.0
2:00 169.5 F 2.0
4:00 198.4 F 5.4
6:00 214.7 F 10.2
8:00 213.1 F 5.3
10:00 236.9 F 9.7
12:00 232.5 F 8.6
14:00 244.2 F 12.4
16:00 241.3 F 11.0
18:00 254.4 F 17.6
20:00 265.8 F 24.5
22:00 273.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.1
1:00 272.4 F 29.0
2:00 272.4 F 29.0
3:00 272.4 F 29.1
4:00 272.2 F 29.3

Min 271.8 F 29.0
Max 272.5 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.7 F 0.0
5:00 226.7 F 0.0
10:00 227.7 F 0.0
15:00 229.9 F 0.0
20:00 218.8 F 0.0
25:00 219.3 F 0.0
30:00 220.3 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 22:08
STERILIZING: 4:00
VENTING: 3:01
DRYING: 30:00
TOTAL CYCLE: 01:00:37
CYCLE COMPLETE

12/01/2021 03:34 PM
ACCEPTED: X
REJECTED:
Operator: U02