

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1340

Name: RITTER2  
Sterilizer ID: V2313228  
07/25/2022 02:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.7 F 0.0  
2:00 159.0 F 0.1  
4:00 186.4 F 2.5  
6:00 205.7 F 7.0  
8:00 221.4 F 12.0  
10:00 228.6 F 7.9  
12:00 239.7 F 13.3  
14:00 240.0 F 11.4  
16:00 239.3 F 10.5  
18:00 252.5 F 17.1  
20:00 264.0 F 23.9  
22:00 271.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.1  
1:00 272.7 F 29.8  
2:00 272.1 F 29.2  
3:00 271.6 F 29.3  
4:00 272.8 F 29.9

Min 271.5 F 28.9  
Max 272.8 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 224.6 F 0.0  
10:00 220.9 F 0.0  
15:00 223.1 F 0.0  
20:00 222.1 F 0.0  
25:00 222.4 F 0.0  
30:00 223.6 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 22:06  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 01:00:25  
CYCLE COMPLETE

07/25/2022 03:45 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_