
Midmark M11 - v1.0.0
Total Cycles: 4397

Name: RITTER 1
Sterilizer ID: V2178198
03/22/2024 11:02 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 150.5 F 1.8
4:00 153.3 F 5.1
6:00 185.5 F 8.7
8:00 202.2 F 13.6
10:00 214.8 F 6.0
12:00 226.0 F 10.5
14:00 230.1 F 7.2
16:00 240.5 F 11.9
18:00 239.1 F 9.9
20:00 251.5 F 16.1
22:00 262.6 F 22.4
24:00 272.8 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.6 F 28.9
2:00 271.3 F 28.4
3:00 272.6 F 29.0
4:00 271.4 F 28.3

Min 271.3 F 28.3
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 231.5 F 0.0
10:00 232.8 F 0.0
15:00 234.1 F 0.0
20:00 223.2 F 0.0
25:00 219.8 F 0.0
30:00 218.9 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:03:03
CYCLE COMPLETE

03/22/2024 12:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____