
Midmark M11 - v1.0.5
Total Cycles: 960

Name: RITTER2
Sterilizer ID: V2313228
04/06/2022 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 96.0 F 0.0
2:00 97.1 F 0.0
4:00 109.5 F 0.0
6:00 146.3 F 0.3
8:00 182.4 F 0.7
10:00 200.4 F 4.0
12:00 214.3 F 8.2
14:00 227.3 F 12.5
16:00 229.7 F 8.4
18:00 241.5 F 13.0
20:00 241.3 F 11.8
22:00 240.8 F 11.0
24:00 253.9 F 17.8
26:00 265.5 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.8
1:00 272.3 F 29.5
2:00 272.2 F 29.0
3:00 271.9 F 29.3
4:00 272.5 F 29.3

Min 271.5 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.7 F 0.0
5:00 219.7 F 0.0
10:00 199.7 F 0.0
15:00 192.9 F 0.0
20:00 201.7 F 0.0
25:00 218.3 F 0.0
30:00 230.5 F 0.0

Drying Complete

FILLING: 0:50
HEATING: 27:37
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:05:35
CYCLE COMPLETE

04/06/2022 10:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____