

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5058

Name: RITTER 1  
Sterilizer ID: V2178198  
11/05/2024 01:42 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.7 F 0.0  
2:00 169.5 F 1.0  
4:00 183.6 F 5.1  
6:00 204.0 F 9.7  
8:00 218.5 F 7.6  
10:00 231.1 F 9.0  
12:00 236.5 F 10.5  
14:00 241.7 F 11.5  
16:00 238.5 F 9.6  
18:00 251.8 F 16.1  
20:00 263.1 F 22.7  
22:00 272.2 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 271.9 F 28.5  
2:00 272.1 F 28.9  
3:00 271.4 F 28.2  
4:00 272.1 F 28.9

Min 271.0 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.5 F 0.0  
5:00 225.5 F 0.0  
10:00 197.5 F 0.0  
15:00 194.4 F 0.0  
20:00 192.7 F 0.0  
25:00 191.2 F 0.0  
30:00 189.6 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 22:11  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:00:48  
CYCLE COMPLETE

11/05/2024 02:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_