
Midmark M11 - v1.0.0
Total Cycles: 667

Name: MIDMARK
Sterilizer ID: V2178198
02/11/2021 04:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.1 F 0.0
2:00 173.1 F 2.2
4:00 176.8 F 5.8
6:00 206.6 F 10.7
8:00 217.9 F 5.6
10:00 230.0 F 10.9
12:00 231.5 F 7.7
14:00 242.5 F 11.5
16:00 245.3 F 13.0
18:00 258.3 F 19.8
20:00 269.6 F 27.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.0
2:00 271.1 F 28.3
3:00 272.4 F 29.0
4:00 271.4 F 28.4

Min 271.1 F 28.2
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.6 F 0.0
5:00 229.8 F 0.0
10:00 231.9 F 0.0
15:00 236.3 F 0.0
20:00 223.1 F 0.0
25:00 213.9 F 0.0
30:00 212.0 F 0.0
Drying Complete

FILLING: 1:36
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:44
CYCLE COMPLETE

02/11/2021 05:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____