
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000901
12:30 03/02/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 3:02
STERILIZING 4:15
136.0 C 228kPa 4:15
136.4 C 232kPa 4:45
136.3 C 231kPa 5:15
136.3 C 232kPa 5:45
136.2 C 231kPa 6:15
136.4 C 233kPa 6:45
136.4 C 232kPa 7:15
136.4 C 230kPa 7:45
136.4 C 230kPa 7:45
VENTING 7:45
AIR DRYING 8:37
CYCLE COMPLETE 23:37

Digital Signature #
2D08E435FF5A49B6
