

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3177

Name: RITTER 1  
Sterilizer ID: V2178198  
04/28/2023 12:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 185.1 F 0.0  
2:00 186.1 F 3.6  
4:00 198.7 F 8.2  
6:00 218.6 F 11.3  
8:00 228.3 F 8.6  
10:00 238.5 F 11.3  
12:00 240.3 F 11.6  
14:00 239.1 F 10.0  
16:00 252.5 F 16.8  
18:00 264.2 F 23.6  
20:00 271.7 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.6  
1:00 272.7 F 29.1  
2:00 271.7 F 28.7  
3:00 272.7 F 29.1  
4:00 271.6 F 28.7  
  
Min 271.6 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.1 F 0.0  
5:00 205.7 F 0.0  
10:00 222.6 F 0.0  
15:00 219.3 F 0.0  
20:00 218.9 F 0.0  
25:00 219.7 F 0.0  
30:00 216.4 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:00  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:49  
CYCLE COMPLETE

04/28/2023 01:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_