
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007127
11:42 27/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:21
STERILIZING 3:24
136.5 C 222kPa 3:24
136.3 C 224kPa 3:54
136.4 C 225kPa 4:24
136.4 C 224kPa 4:54
136.4 C 224kPa 5:24
136.1 C 222kPa 5:54
136.2 C 218kPa 6:24
136.4 C 223kPa 6:54
136.4 C 223kPa 6:54
VENTING 6:55
AIR DRYING 7:45
CYCLE COMPLETE 37:45

Digital Signature #
AE845697719DEE65
