
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000177
8:47 06/07/2020

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:06
PRESSURIZING 4:08
STERILIZING 5:25
136.4 C 224kPa 5:25
136.4 C 221kPa 5:55
136.1 C 221kPa 6:25
136.3 C 223kPa 6:55
136.4 C 222kPa 7:25
136.4 C 224kPa 7:55
136.2 C 219kPa 8:25
136.3 C 222kPa 8:55
136.3 C 222kPa 8:55
VENTING 8:55
AIR DRYING 9:50
CYCLE COMPLETE 9:55

Digital Signature #
4CBA1E2DD8D720A3
