
Midmark M11 - v1.0.5
Total Cycles: 2837

Name: RITTER2
Sterilizer ID: V2313228
09/06/2023 11:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.4 F 0.0
2:00 143.2 F 0.0
4:00 169.6 F 2.8
6:00 195.0 F 7.1
8:00 209.4 F 12.2
10:00 222.7 F 7.3
12:00 236.4 F 12.5
14:00 235.7 F 9.8
16:00 237.7 F 9.1
18:00 248.1 F 14.9
20:00 260.1 F 21.6
22:00 270.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.0
1:00 272.4 F 29.6
2:00 271.2 F 29.1
3:00 272.4 F 29.7
4:00 271.5 F 29.0

Min 271.2 F 28.9
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 203.1 F 0.0
10:00 213.5 F 0.0
15:00 210.8 F 0.0
20:00 210.0 F 0.0
25:00 212.3 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 22:43
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:54
CYCLE COMPLETE

09/06/2023 12:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____