
Midmark M11 - v1.0.0
Total Cycles: 5105

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2024 05:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 129.9 F 0.0
2:00 132.4 F 0.0
4:00 157.7 F 0.3
6:00 183.6 F 0.8
8:00 205.0 F 4.5
10:00 218.2 F 8.3
12:00 228.5 F 12.7
14:00 230.5 F 8.3
16:00 243.4 F 13.5
18:00 241.5 F 11.4
20:00 238.9 F 9.9
22:00 252.5 F 16.5
24:00 263.7 F 23.0
26:00 271.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.7 F 29.0
2:00 271.8 F 28.3
3:00 271.8 F 28.5
4:00 272.6 F 28.9

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.7 F 0.0
5:00 196.8 F 0.0
10:00 199.2 F 0.0
15:00 199.8 F 0.0
20:00 196.4 F 0.0
25:00 189.9 F 0.0
30:00 186.4 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 26:09
STERILIZING: 4:00
VENTING: 3:35
DRYING: 30:00
TOTAL CYCLE: 01:04:37
CYCLE COMPLETE

11/18/2024 06:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____