

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2930

Name: RITTER2  
Sterilizer ID: V2313228  
09/26/2023 12:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 193.9 F 0.0  
2:00 187.4 F 3.7  
4:00 202.6 F 8.8  
6:00 218.9 F 10.8  
8:00 225.3 F 9.4  
10:00 234.9 F 9.7  
12:00 241.8 F 12.8  
14:00 242.5 F 11.9  
16:00 254.8 F 18.3  
18:00 265.6 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.9  
1:00 272.4 F 29.6  
2:00 271.5 F 28.9  
3:00 271.8 F 29.6  
4:00 272.0 F 29.4  
  
Min 271.0 F 28.8  
Max 272.5 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.9 F 0.0  
5:00 222.6 F 0.0  
10:00 228.1 F 0.0  
15:00 229.0 F 0.0  
20:00 229.8 F 0.0  
25:00 227.3 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 1:56  
HEATING: 19:46  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

09/26/2023 01:30 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_