

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3013

Name: RITTER 1  
Sterilizer ID: V2178198  
03/16/2023 11:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.2 F 0.0  
2:00 187.2 F 3.6  
4:00 200.1 F 7.8  
6:00 214.7 F 13.0  
8:00 226.3 F 7.7  
10:00 238.6 F 12.8  
12:00 237.7 F 10.0  
14:00 237.1 F 8.4  
16:00 250.0 F 15.1  
18:00 262.1 F 22.0  
20:00 272.7 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.2  
1:00 272.6 F 28.9  
2:00 271.6 F 28.2  
3:00 272.4 F 29.0  
4:00 272.3 F 28.7  
  
Min 271.3 F 28.1  
Max 272.7 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
5:00 216.8 F 0.0  
10:00 222.6 F 0.0  
15:00 223.5 F 0.0  
20:00 229.6 F 0.0  
25:00 221.3 F 0.0  
30:00 216.1 F 0.0  
Drying Complete

FILLING: 1:34  
HEATING: 20:21  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:08  
CYCLE COMPLETE

03/16/2023 12:20 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_