
Midmark M11 - v1.0.5
Total Cycles: 1498

Name: RITTER2
Sterilizer ID: V2313228
09/19/2022 07:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.7 F 0.0
2:00 75.3 F 0.0
4:00 99.8 F 0.0
6:00 133.2 F 0.6
8:00 159.0 F 3.5
10:00 175.0 F 6.8
12:00 198.0 F 10.5
14:00 216.5 F 6.2
16:00 223.0 F 7.6
18:00 231.5 F 11.7
20:00 231.8 F 8.0
22:00 241.8 F 13.2
24:00 241.2 F 11.5
26:00 253.4 F 17.7
28:00 263.7 F 24.1
30:00 272.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.4
1:00 272.4 F 29.7
2:00 272.3 F 30.1
3:00 271.7 F 29.3
4:00 272.4 F 29.9

Min 271.6 F 29.3
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 202.6 F 0.0
10:00 210.8 F 0.0
15:00 215.1 F 0.0
20:00 218.6 F 0.0
25:00 225.0 F 0.0
30:00 230.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:06
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:08:05
CYCLE COMPLETE

09/19/2022 08:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____