
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000904
12:52 30/11/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:46
PRESSURIZING 3:46
STERILIZING 4:49
136.4 C 225kPa 4:49
136.2 C 224kPa 5:19
136.3 C 223kPa 5:49
136.3 C 223kPa 6:19
136.4 C 223kPa 6:49
136.0 C 222kPa 7:19
136.4 C 224kPa 7:49
136.4 C 224kPa 8:19
136.4 C 224kPa 8:19
VENTING 8:19
AIR DRYING 9:14
CYCLE COMPLETE 9:16

Digital Signature #
BD8518DD297D5C39
