
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002575
10:18 29/10/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:40
PRESSURIZING 2:43
STERILIZING 3:43
136.0 C 220kPa 3:43
136.3 C 221kPa 4:13
136.4 C 223kPa 4:43
136.4 C 222kPa 5:13
136.3 C 223kPa 5:43
135.9 C 220kPa 6:13
136.4 C 222kPa 6:43
136.4 C 224kPa 7:13
136.4 C 224kPa 7:13
VENTING 7:13
AIR DRYING 8:07
CYCLE COMPLETE 8:23

Digital Signature #
99C66D2E9F5627B1
