
Midmark M11 - v1.0.0
Total Cycles: 2431

Name: RITTER 1
Sterilizer ID: V2178198
10/14/2022 09:12 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.9 F 0.0
2:00 165.7 F 2.0
4:00 178.7 F 5.4
6:00 191.1 F 9.4
8:00 201.6 F 8.0
10:00 222.7 F 6.5
12:00 233.1 F 11.4
14:00 232.0 F 7.6
16:00 244.5 F 13.4
18:00 242.5 F 11.6
20:00 254.7 F 18.0
22:00 265.7 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.3 F 29.0
2:00 271.7 F 28.8
3:00 272.1 F 28.9
4:00 271.7 F 28.9

Min 271.5 F 28.5
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.0 F 0.0
5:00 222.5 F 0.0
10:00 237.3 F 0.0
15:00 232.7 F 0.0
20:00 221.4 F 0.0
25:00 217.7 F 0.0
30:00 216.5 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 23:42
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:46
CYCLE COMPLETE

10/14/2022 10:13 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____