

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3320

Name: RITTER2  
Sterilizer ID: V2313228  
01/24/2024 06:39 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.4 F 0.0  
2:00 178.9 F 3.4  
4:00 201.6 F 8.8  
6:00 220.7 F 9.2  
8:00 232.1 F 9.9  
10:00 234.5 F 8.4  
12:00 241.5 F 11.0  
14:00 248.3 F 14.6  
16:00 261.3 F 21.9  
18:00 272.6 F 29.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.7  
1:00 272.1 F 29.0  
2:00 272.3 F 29.5  
3:00 271.4 F 28.7  
4:00 272.3 F 29.4  
  
Min 271.0 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 179.1 F 0.0  
5:00 236.5 F 0.0  
10:00 206.3 F 0.0  
15:00 191.9 F 0.0  
20:00 189.1 F 0.0  
25:00 187.9 F 0.0  
30:00 186.7 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 18:19  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:57:39  
CYCLE COMPLETE

01/24/2024 07:36 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_