

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1739

Name: RITTER 1  
Sterilizer ID: V2178198  
03/24/2022 09:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
2:00 170.6 F 2.4  
4:00 183.8 F 6.0  
6:00 217.5 F 11.3  
8:00 223.4 F 6.2  
10:00 238.7 F 11.4  
12:00 234.0 F 8.4  
14:00 241.0 F 10.7  
16:00 246.2 F 13.3  
18:00 258.5 F 20.1  
20:00 269.9 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.2 F 28.7  
2:00 272.2 F 29.1  
3:00 271.7 F 28.4  
4:00 272.5 F 29.1  
  
Min 271.1 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.5 F 0.0  
5:00 235.2 F 0.0  
10:00 224.2 F 0.0  
15:00 229.1 F 0.0  
20:00 227.1 F 0.0  
25:00 220.8 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 1:20  
HEATING: 20:53  
STERILIZING: 4:00  
VENTING: 3:06  
DRYING: 30:00  
TOTAL CYCLE: 00:59:19  
CYCLE COMPLETE

03/24/2022 10:27 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02