
Midmark M11 - v1.0.5
Total Cycles: 3004

Name: RITTER2
Sterilizer ID: V2313228
10/12/2023 02:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.7 F 0.0
2:00 174.2 F 4.0
4:00 191.1 F 8.9
6:00 212.1 F 7.8
8:00 227.5 F 9.1
10:00 234.4 F 9.5
12:00 243.0 F 13.2
14:00 243.7 F 12.7
16:00 257.2 F 19.9
18:00 269.1 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.8 F 29.9
2:00 271.9 F 29.1
3:00 272.1 F 29.8
4:00 272.8 F 29.8

Min 271.5 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 219.2 F 0.0
10:00 221.9 F 0.0
15:00 212.4 F 0.0
20:00 214.8 F 0.0
25:00 219.9 F 0.0
30:00 219.4 F 0.0
Drying Complete

FILLING: 1:30
HEATING: 19:03
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:47
CYCLE COMPLETE

10/12/2023 03:38 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____