

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2485

Name: RITTER 1  
Sterilizer ID: V2178198  
10/25/2022 11:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
2:00 199.6 F 4.4  
4:00 200.1 F 9.6  
6:00 219.1 F 6.3  
8:00 228.7 F 10.0  
10:00 232.0 F 7.9  
12:00 242.9 F 12.9  
14:00 242.2 F 11.4  
16:00 255.3 F 18.2  
18:00 266.6 F 25.1  
20:00 273.1 F 29.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.0  
1:00 272.7 F 29.4  
2:00 271.9 F 29.3  
3:00 272.5 F 29.4  
4:00 272.0 F 29.4

Min 271.8 F 28.9  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 220.4 F 0.0  
10:00 230.7 F 0.0  
15:00 225.7 F 0.0  
20:00 222.1 F 0.0  
25:00 220.0 F 0.0  
30:00 218.6 F 0.0  
Drying Complete

FILLING: 2:16  
HEATING: 20:20  
STERILIZING: 4:00  
VENTING: 3:13  
DRYING: 30:00  
TOTAL CYCLE: 00:59:49  
CYCLE COMPLETE

10/25/2022 12:04 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_