
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001617
16:00 20/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:14
PRESSURIZING 3:15
STERILIZING 4:13
136.4 C 225kPa 4:13
136.3 C 225kPa 4:43
136.4 C 226kPa 5:13
136.2 C 225kPa 5:43
136.3 C 224kPa 6:13
136.3 C 225kPa 6:43
136.4 C 226kPa 7:13
136.4 C 226kPa 7:43
136.4 C 226kPa 7:43
VENTING 7:43
AIR DRYING 8:42
CYCLE COMPLETE 8:48

Digital Signature #
43F65623A42F3B19
