

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4893

Name: RITTER 1  
Sterilizer ID: V2178198  
09/11/2024 02:19 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 155.2 F 0.0  
2:00 174.6 F 0.4  
4:00 185.3 F 1.5  
6:00 205.1 F 5.4  
8:00 220.0 F 9.5  
10:00 229.1 F 11.2  
12:00 233.7 F 9.2  
14:00 239.9 F 10.9  
16:00 243.8 F 12.3  
18:00 241.9 F 11.0  
20:00 254.1 F 17.3  
22:00 264.9 F 23.7  
24:00 271.7 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.4 F 28.7  
2:00 272.6 F 28.7  
3:00 272.0 F 28.2  
4:00 271.5 F 28.0

Min 271.4 F 27.9  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 211.4 F 0.0  
5:00 198.5 F 0.0  
10:00 208.9 F 0.0  
15:00 199.5 F 0.0  
20:00 180.7 F 0.0  
25:00 198.2 F 0.0  
30:00 205.9 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 24:07  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:02:54  
CYCLE COMPLETE

09/11/2024 03:21 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_