
Midmark M11 - v1.0.5
Total Cycles: 2502

Name: RITTER2
Sterilizer ID: V2313228
06/12/2023 02:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.5 F 0.0
2:00 181.6 F 3.4
4:00 196.0 F 8.2
6:00 216.4 F 10.6
8:00 229.3 F 9.3
10:00 233.9 F 9.3
12:00 242.0 F 12.7
14:00 241.6 F 11.7
16:00 254.6 F 18.4
18:00 265.8 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.0 F 29.4
2:00 272.5 F 29.7
3:00 272.0 F 29.9
4:00 271.7 F 29.3

Min 271.4 F 29.2
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 218.5 F 0.0
10:00 247.7 F 0.0
15:00 237.4 F 0.0
20:00 223.8 F 0.0
25:00 225.3 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 19:39
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

06/12/2023 03:06 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____