
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001725
13:40 27/03/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:25
PRESSURIZING 2:33
STERILIZING 3:31
136.3 C 220kPa 3:31
136.2 C 221kPa 4:01
136.3 C 224kPa 4:31
136.2 C 225kPa 5:01
136.1 C 222kPa 5:31
136.3 C 225kPa 6:01
136.4 C 226kPa 6:31
136.4 C 225kPa 7:01
136.4 C 225kPa 7:01
VENTING 7:01
AIR DRYING 7:54
CYCLE COMPLETE 8:00

Digital Signature #
7184A6C2D2BA4800
