
Surrey, BC
Canada

STATIM 5000 S501R711
SN 511018F00076
Unit # : 000
WATER QUALITY
2.4uS / 1.5ppm
CYCLE NUMBER 000013
13:39 13/01/2020
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:15
PRESSURIZING 6:16
STERILIZING 8:10
136.3 C 226kPa 8:10
136.4 C 227kPa 8:40
136.2 C 226kPa 9:10
136.4 C 227kPa 9:40
136.4 C 228kPa 10:10
136.4 C 227kPa 10:40
136.4 C 227kPa 11:11
136.4 C 227kPa 11:40
VENTING 11:41
AIR DRYING 12:57
CYCLE COMPLETE 42:58

Digital Signature #
EE3E37729CD590C0
