
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005160
16:39 27/02/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:57
PRESSURIZING 3:00
STERILIZING 4:02
136.4 C 222kPa 4:02
136.3 C 225kPa 4:32
136.3 C 223kPa 5:02
136.2 C 223kPa 5:32
136.4 C 225kPa 6:02
136.1 C 218kPa 6:32
136.4 C 223kPa 7:02
136.3 C 223kPa 7:32
136.3 C 223kPa 7:32
VENTING 7:32
AIR DRYING 8:22
CYCLE COMPLETE 10:05

Digital Signature #
46E9F4473D7EFBA4
