
Midmark M11 - v1.0.5
Total Cycles: 3290

Name: RITTER2
Sterilizer ID: V2313228
01/11/2024 01:51 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.1 F 0.0
2:00 163.7 F 2.5
4:00 183.1 F 6.2
6:00 201.2 F 10.4
8:00 224.3 F 6.2
10:00 228.6 F 9.6
12:00 235.9 F 9.6
14:00 242.9 F 12.5
16:00 241.9 F 11.7
18:00 254.7 F 18.2
20:00 265.8 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.5 F 29.6
2:00 271.1 F 28.8
3:00 272.6 F 30.1
4:00 271.6 F 29.4

Min 271.0 F 28.8
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.5 F 0.0
5:00 233.9 F 0.0
10:00 226.0 F 0.0
15:00 193.1 F 0.0
20:00 183.4 F 0.0
25:00 181.6 F 0.0
30:00 180.7 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

01/11/2024 02:51 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____