
Midmark M11 - v1.0.0
Total Cycles: 3897

Name: RITTER 1
Sterilizer ID: V2178198
10/12/2023 01:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.3 F 0.0
2:00 163.8 F 3.4
4:00 184.0 F 7.8
6:00 225.5 F 13.2
8:00 226.3 F 7.4
10:00 237.1 F 12.9
12:00 237.7 F 10.0
14:00 235.8 F 8.3
16:00 249.9 F 15.1
18:00 262.2 F 22.0
20:00 272.7 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.4 F 28.7
2:00 273.0 F 29.1
3:00 271.7 F 28.8
4:00 272.3 F 28.7

Min 271.5 F 28.4
Max 273.0 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.2 F 0.0
5:00 238.9 F 0.0
10:00 228.5 F 0.0
15:00 224.2 F 0.0
20:00 212.2 F 0.0
25:00 187.7 F 0.0
30:00 180.1 F 0.0
Drying Complete

FILLING: 1:25
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:22
DRYING: 30:00
TOTAL CYCLE: 00:59:43
CYCLE COMPLETE

10/12/2023 02:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____