

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1177

Name: RITTER 1  
Sterilizer ID: V2178198  
09/16/2021 07:02 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 72.6 F 0.0  
2:00 73.8 F 0.0  
4:00 78.2 F 0.0  
6:00 101.6 F 0.0  
8:00 139.4 F 0.1  
10:00 168.0 F 0.1  
12:00 195.3 F 0.2  
14:00 214.8 F 2.2  
16:00 223.3 F 4.8  
18:00 229.2 F 7.2  
20:00 236.5 F 10.3  
22:00 240.2 F 10.3  
24:00 241.6 F 11.2  
26:00 236.7 F 8.7  
28:00 242.0 F 10.5  
30:00 247.2 F 13.4  
32:00 258.3 F 19.4  
34:00 268.3 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 27.8  
1:00 272.5 F 28.6  
2:00 271.7 F 27.8  
3:00 271.9 F 28.4  
4:00 272.3 F 28.3

Min 271.2 F 27.7  
Max 272.6 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 177.3 F 0.0  
5:00 218.6 F 0.0  
10:00 204.8 F 0.0  
15:00 202.0 F 0.0  
20:00 185.9 F 0.0  
25:00 197.8 F 0.0  
30:00 202.8 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 35:20  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 01:13:25  
CYCLE COMPLETE

09/16/2021 08:19 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02