
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001219
9:25 18/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:50
PRESSURIZING 3:52
STERILIZING 5:11
136.0 C 220kPa 5:11
136.4 C 223kPa 5:41
136.3 C 223kPa 6:11
136.4 C 222kPa 6:41
136.3 C 223kPa 7:11
136.4 C 224kPa 7:41
136.4 C 221kPa 8:11
136.4 C 225kPa 8:41
136.4 C 225kPa 8:41
VENTING 8:42
AIR DRYING 9:39
CYCLE COMPLETE 14:11

Digital Signature #
E1ED82F3D380F031
