
Midmark M11 - v1.0.5
Total Cycles: 109

Name: RITTER2
Sterilizer ID: V2313228
06/16/2021 02:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 168.9 F 0.0
2:00 170.7 F 2.4
4:00 174.4 F 6.2
6:00 195.6 F 10.6
8:00 214.0 F 6.1
10:00 225.2 F 9.8
12:00 232.2 F 9.7
14:00 241.3 F 13.0
16:00 242.4 F 12.3
18:00 254.1 F 18.3
20:00 264.7 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.6
1:00 271.6 F 29.5
2:00 272.1 F 29.7
3:00 271.9 F 29.9
4:00 271.8 F 30.1

Min 271.3 F 29.5
Max 272.9 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.3 F 0.0
5:00 224.0 F 0.0
10:00 233.5 F 0.0
15:00 234.5 F 0.0
20:00 226.2 F 0.0
25:00 196.0 F 0.0
30:00 176.4 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 21:59
STERILIZING: 4:00
VENTING: 3:33
DRYING: 30:00
TOTAL CYCLE: 01:01:13
CYCLE COMPLETE

06/16/2021 03:24 PM
ACCEPTED: X
REJECTED:
Operator: U01