

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5330

Name: RITTER 1  
Sterilizer ID: V2178198  
02/10/2025 03:35 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 140.5 F 0.0  
2:00 138.5 F 2.1  
4:00 163.6 F 5.3  
6:00 182.5 F 9.2  
8:00 230.1 F 11.9  
10:00 225.0 F 5.9  
12:00 233.5 F 10.2  
14:00 234.5 F 7.9  
16:00 241.0 F 11.0  
18:00 236.1 F 8.5  
20:00 249.3 F 14.6  
22:00 260.1 F 20.6  
24:00 270.0 F 26.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 27.9  
1:00 271.4 F 28.5  
2:00 272.5 F 29.0  
3:00 271.3 F 28.4  
4:00 272.3 F 29.1

Min 270.6 F 27.8  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.1 F 0.0  
5:00 229.1 F 0.0  
10:00 206.4 F 0.0  
15:00 229.1 F 0.0  
20:00 207.5 F 0.0  
25:00 185.4 F 0.0  
30:00 184.8 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 24:55  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:03:44  
CYCLE COMPLETE

02/10/2025 04:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_