
Midmark M11 - v1.0.5
Total Cycles: 2389

Name: RITTER2
Sterilizer ID: V2313228
05/18/2023 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.3 F 0.0
2:00 143.2 F 0.2
4:00 164.2 F 2.9
6:00 189.1 F 7.3
8:00 209.6 F 12.2
10:00 222.3 F 6.3
12:00 235.1 F 11.5
14:00 237.0 F 9.5
16:00 237.7 F 9.2
18:00 250.4 F 15.8
20:00 262.6 F 23.0
22:00 272.0 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.8
1:00 272.2 F 29.8
2:00 272.7 F 29.5
3:00 272.7 F 30.1
4:00 272.7 F 29.6

Min 271.0 F 28.7
Max 273.3 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 219.5 F 0.0
10:00 222.2 F 0.0
15:00 205.5 F 0.0
20:00 204.8 F 0.0
25:00 234.9 F 0.0
30:00 230.4 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 22:02
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

05/18/2023 12:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____