
Midmark M11 - v1.0.0
Total Cycles: 3958

Name: RITTER 1
Sterilizer ID: V2178198
10/30/2023 01:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.1 F 0.0
2:00 152.0 F 2.7
4:00 172.9 F 6.5
6:00 220.1 F 11.8
8:00 225.0 F 5.9
10:00 230.0 F 10.9
12:00 231.5 F 7.3
14:00 245.1 F 13.5
16:00 244.2 F 12.3
18:00 257.2 F 19.1
20:00 268.7 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.6
1:00 272.1 F 28.6
2:00 272.1 F 28.6
3:00 272.5 F 28.8
4:00 272.7 F 29.0

Min 271.4 F 28.5
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.1 F 0.0
5:00 213.1 F 0.0
10:00 199.8 F 0.0
15:00 191.9 F 0.0
20:00 189.4 F 0.0
25:00 187.2 F 0.0
30:00 177.8 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 21:32
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:31
CYCLE COMPLETE

10/30/2023 02:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____