
Midmark M11 - v1.0.5
Total Cycles: 1035

Name: RITTER2
Sterilizer ID: V2313228
04/26/2022 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 135.8 F 0.0
2:00 136.1 F 2.0
4:00 162.6 F 5.6
6:00 187.1 F 10.1
8:00 211.4 F 5.3
10:00 225.8 F 8.3
12:00 237.8 F 13.5
14:00 237.5 F 10.6
16:00 235.0 F 8.8
18:00 249.5 F 15.6
20:00 261.3 F 22.3
22:00 272.0 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.4 F 29.9
2:00 272.3 F 29.3
3:00 271.5 F 29.1
4:00 272.5 F 29.9

Min 271.2 F 28.9
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.9 F 0.0
5:00 228.9 F 0.0
10:00 252.0 F 0.0
15:00 226.4 F 0.0
20:00 225.6 F 0.0
25:00 222.8 F 0.0
30:00 213.4 F 0.0
Drying Complete

FILLING: 1:08
HEATING: 22:36
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:00:56
CYCLE COMPLETE

04/26/2022 01:24 PM
ACCEPTED: X
REJECTED:
Operator: U02