

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3102

Name: RITTER2  
Sterilizer ID: V2313228  
11/08/2023 04:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 143.9 F 0.0  
2:00 144.7 F 1.7  
4:00 163.0 F 5.4  
6:00 191.9 F 10.5  
8:00 212.2 F 4.4  
10:00 224.0 F 10.4  
12:00 229.7 F 7.4  
14:00 240.9 F 11.8  
16:00 244.8 F 13.3  
18:00 258.1 F 20.5  
20:00 270.2 F 28.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 29.0  
1:00 272.3 F 30.0  
2:00 272.7 F 29.7  
3:00 271.7 F 29.1  
4:00 271.3 F 29.2  
  
Min 271.2 F 28.9  
Max 272.7 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
5:00 229.9 F 0.0  
10:00 225.2 F 0.0  
15:00 227.5 F 0.0  
20:00 214.1 F 0.0  
25:00 203.8 F 0.0  
30:00 198.7 F 0.0  
Drying Complete

FILLING: 1:08  
HEATING: 20:58  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:18  
CYCLE COMPLETE

11/08/2023 05:28 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_