

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000130  
13:10 08/09/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 4:16  
PRESSURIZING 4:16  
STERILIZING 5:44  
136.0 C 221kPa 5:44  
136.3 C 225kPa 6:14  
136.4 C 225kPa 6:44  
136.4 C 225kPa 7:14  
136.4 C 223kPa 7:44  
136.4 C 225kPa 8:14  
136.3 C 223kPa 8:44  
136.1 C 222kPa 9:14  
136.1 C 222kPa 9:14  
VENTING 9:14  
AIR DRYING 10:15  
CYCLE COMPLETE 15:00  
-----

Digital Signature #  
EA88712CAF3D5C7F  
-----