
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000005
16:13 07/01/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:53
PRESSURIZING 2:57
STERILIZING 4:01
136.0 C 220kPa 4:01
135.7 C 221kPa 4:31
136.3 C 223kPa 5:01
136.4 C 223kPa 5:31
136.1 C 221kPa 6:01
136.4 C 224kPa 6:31
136.2 C 224kPa 7:01
136.3 C 222kPa 7:31
136.3 C 222kPa 7:31
VENTING 7:31
AIR DRYING 8:24
CYCLE COMPLETE 8:31

Digital Signature #
C48022DF3CA00003
