
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000178
9:28 06/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:58
PRESSURIZING 4:00
STERILIZING 5:04
136.0 C 220kPa 5:04
136.1 C 221kPa 5:34
136.4 C 222kPa 6:04
136.4 C 222kPa 6:34
136.4 C 223kPa 7:04
CYCLE ABORTED 7:16
STOP BUTTON PRESSED
NOT STERILE
END TIME 7:16

Digital Signature #
CBF5789380C750ED
