
Midmark M11 - v1.0.0
Total Cycles: 2330

Name: RITTER 1
Sterilizer ID: V2178198
09/15/2022 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.3 F 0.0
2:00 189.2 F 3.1
4:00 204.1 F 7.7
6:00 222.3 F 11.7
8:00 228.3 F 8.0
10:00 240.3 F 13.2
12:00 238.7 F 10.5
14:00 237.8 F 8.5
16:00 250.5 F 15.6
18:00 262.8 F 22.5
20:00 272.1 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.3 F 28.8
2:00 271.7 F 28.8
3:00 272.1 F 28.6
4:00 271.6 F 28.7

Min 271.4 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.9 F 0.0
5:00 226.1 F 0.0
10:00 237.7 F 0.0
15:00 228.4 F 0.0
20:00 225.7 F 0.0
25:00 225.0 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 2:03
HEATING: 20:11
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:25
CYCLE COMPLETE

09/15/2022 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____